

Pt. 53, Subpt. C, Table C-2

40 CFR Ch. I (7-1-11 Edition)

Pollutant	Concentration range, parts per million (ppm)	Simultaneous measurements required				Maximum discrepancy specification, parts per million
		1-hour		24-hour		
		First set	Second set	First set	Second set	
	Med. 0.10 to 0.20 .....	.....	.....	2	2	0.02
	High 0.25 .....	.....	.....	2	2	0.03
	Total .....	.....	.....	7	8	.....

[75 FR 35601, June 22, 2010]

TABLE C-2 TO SUBPART C OF PART 53—SEQUENCE OF TEST MEASUREMENTS

Measurement	Concentration range	
	First set	Second set
1 .....	Low .....	Medium.
2 .....	High .....	High.
3 .....	Medium .....	Low.
4 .....	High .....	High.
5 .....	Low .....	Medium.
6 .....	Medium .....	Low.
7 .....	Low .....	Medium.
8 .....	Medium .....	Low.
9 .....	High .....	High.
10 .....	Medium .....	Low.
11 .....	High .....	Medium.
12 .....	Low .....	High.
13 .....	Medium .....	Medium.
14 .....	Low .....	High.
15 .....	.....	Low.
16 .....	.....	Medium.
17 .....	.....	Low.
18 .....	.....	High.

TABLE C-3 TO SUBPART C OF PART 53—TEST SPECIFICATIONS FOR PB IN TSP AND PB IN PM<sub>10</sub> METHODS

TABLE C-3 TO SUBPART C OF PART 53—TEST SPECIFICATIONS FOR PB IN TSP AND PB IN PM<sub>10</sub> METHODS

Concentration range equivalent to percentage of NAAQS in $\mu\text{g}/\text{m}^3$ .	30% to 250%
Minimum number of 24-hr measurements.	5
Maximum reference method analytical bias, $D_q$ .	$\pm 5\%$
Maximum precision, $P_R$ or $P_C$ .....	$\leq 15\%$
Maximum difference (D) .....	$\pm 20\%$
Estimated Method Detection Limit (MDL), $\mu\text{g}/\text{m}^3$ .	5% of NAAQS level.

[73 FR 67059, Nov. 12, 2008]

TABLE C-4 TO SUBPART C OF PART 53—TEST SPECIFICATIONS FOR PM<sub>10</sub>, PM<sub>2.5</sub> AND PM<sub>10-2.5</sub> CANDIDATE EQUIVALENT METHODS

Specification	PM <sub>10</sub>	PM <sub>2.5</sub>			PM <sub>10-2.5</sub>	
		Class I	Class II	Class III	Class II	Class III
Acceptable concentration range ( $R_i$ ), $\mu\text{g}/\text{m}^3$ .	15-300	3-200 ..	3-200	3-200	3-200	3-200
Minimum number of test sites.	2 .....	1 .....	2	4	2	4
Minimum number of candidate method samplers or analyzers per site.	3 .....	3 .....	3 <sup>1</sup>	3 <sup>1</sup>	3 <sup>1</sup>	3 <sup>1</sup>
Number of reference method samplers per site.	3 .....	3 .....	3 <sup>1</sup>	3 <sup>1</sup>	3 <sup>1</sup>	3 <sup>1</sup>
Minimum number of acceptable sample sets per site for PM <sub>10</sub> methods:						
$R_i < 60 \mu\text{g}/\text{m}^3$ .....	3					
$R_i > 60 \mu\text{g}/\text{m}^3$ .....	3					
Total .....	10					
Minimum number of acceptable sample sets per site for PM <sub>2.5</sub> and PM <sub>10-2.5</sub> candidate equivalent methods:						
$R_i < 30 \mu\text{g}/\text{m}^3$ for 24-hr or $R_i < 20 \mu\text{g}/\text{m}^3$ for 48-hr samples.	.....	3				